04 TdA-JP4.02 Gramicidin

Document-Id: Document-1997 (Revision: 3)

Document type: Turbidimetric assay (USP <81>, JP 4.02)/1

06.09.2021 11:02:31 Last modified: Database: O130_Permanent_PLA_305 Database GUID: 996faea8-9ec8-4fb7-8524-ea8686ffdf5d

Calculation information

Calculation performed: 06.09.2021 11:02:31, Benjamin Schneider (PLA 3.0.5 Build 816, NBARW01, 10014) 06.09.2021 11:12:55, Benjamin Schneider (PLA 3.0.5 Build 816, NBARW01, 10014) Report generated:

Signatures

Responsibility	Review	Approval

Comments

Comment by Benjamin Schneider: (11.08.2021 16:43:04)

Example turbidimetric assay based on JP 4.02: The standard consists of five concentration steps (S1 to S5) with S3 being used as reference. The relative potency of the unkown sample TST1 is estimated based on the standard. The test system includes tests on the combined standard deviation of the standard, the coefficient of determination (r^2), and the relative potency.



Documentation

Date 11.06.2021 15:44:17

Substance Gramicidin, Polypeptide, Gramicidin is a mixture of peptide substances having antibacterial activity

produced by the growth of Bacillus brevis Dubos. Gramicidin occurs as a white to light yellowish white

crystalline powder.

Inoculum

Test organism

Species Enterococcus hirae

Incubation 22h, 36.5° C

Incubation medium

Ingredient Glucose, 10g
Ingredient Casein peptone, 5g
Ingredient Yeast extract, 20g

Ingredient Potassium dihydrogen phosphate, 2g

Ingredient Polysorbate 80, 0.1g

Ingredient Agar, 15g Ingredient Water, 1000ml

pH 6.7

Inoculum medium

Ingredient Glucose, 1g
Ingredient Casein peptone, 5g
Ingredient Meat extract, 1.5g
Ingredient Yeast extract, 1.5g
Ingredient Sodium chloride, 3.5g

Ingredient Potassium dihydrogen phosphate, 1.32g Ingredient Disodium hydrogen phosphate, 3g

Ingredient Water, 1000ml

pH 7

Inoculum concentration 0.02ug/ml Inoculum volume per tube 10ml

Incubation 250min, 37°C

Assay overview

Assay elements	Summary
■ STD	Standard sample, 3 replicates, Sequence: 0.065, 0.08, 0.1, 0.125, 0.155
■ TST1	Test sample, 3 replicates, Predilution: 1

Property	Value
Response data processing* Response adjustment Response normalization Response transformation Replicate averaging	No adjustment No normalization No transformation Averaging deactivated
Dose transformation Response unit	No dose transformation (not available)

^{*} The response data processing steps are executed in the reported order.

Rack layout

		1		2		3
1	[1]	0.855	[2]	0.814	[3]	0.628
2	[4]	0.693	[5]	0.530	[1]	0.713

	1	2	3
1	[1] 0.842	[2] 0.827	[3] 0.695
2	[4] 0.685	[5] 0.578	[1] 0.796

	1	2	3
1	[1] 0.850	[2] 0.839	[3] 0.756
2	[4] 0.670	[5] 0.532	[1] 0.720

Assay suitability tests

	Passed	Failed (rejected)	Failed (warning)	Info	
Assay suitability tests	2	0	0	0	
Overall test result		Passed			

Test: Combined standard deviation of the standard Scope: Standard only; Severity level: Warning				
Calculated value Upper margin	0.03215 0.10000			
Test: Coefficient of determination (R²), regular Scope: Standard only; Severity level: Warning				
Calculated value Lower margin	0.94385 0.90000			


■ STD: Setup

Property	Value
Concentration unit Number of replicates Stepwise concentration sequence	μg/ml 3 0.065, 0.08, 0.1, 0.125, 0.155
Reference step Position Concentration	3 0.1

■ STD: Result

Data plot

Regression plot



Response data

Dilution step	1	2	3	4	5
Working concentration [µg/ml]	0.06500	0.08000	0.10000	0.12500	0.15500
Response	0.85450	0.81420	0.62840	0.69330	0.52990
	0.84220	0.82730	0.69470	0.68500	0.57790
	0.84950	0.83920	0.75630	0.66990	0.53160
Mean	0.84873	0.82690	0.69313	0.68273	0.54647
SD	0.00619	0.01250	0.06396	0.01186	0.02724
CV	0.73%	1.51%	9.23%	1.74%	4.98%

Response data processing

Processing instructions

 $\textit{Adjustment} \rightarrow \textit{Normalization} \rightarrow \textit{Transformation} \rightarrow \textit{Averaging}$

Step	Details
Response adjustment	No adjustment
Response normalization	No normalization
Response transformation	No transformation
Replicate averaging	Averaging deactivated

Regression

Parameter	Estimate	Error
Intercept	1.06843	0.05137
Slope	-3.32223	0.46785

	Quality of regression	
r²	0.94385	
r² adjusted	0.92513	

Sample suitability tests


	Passed	Failed (rejected)	Failed (warning)	Info
Assay suitability tests	2	0	0	0
Sample suitability tests	0	0	0	0
Overall test result	Passed			

■ TST1: Setup

Property	Value
Concentration unit	μg/ml
Number of replicates	3
Predilution	1

■ TST1: Result

Interpolation plot



Response data

	Raw data
Response	0.71300
	0.79600
	0.72010
Mean	0.74303
SD	0.04601
CV	6.19%

Response data processing

Processing instructions

 $\textit{Adjustment} \rightarrow \textit{Normalization} \rightarrow \textit{Transformation} \rightarrow \textit{Averaging}$

Step	Details
Response adjustment	No adjustment
Response normalization	No normalization
Response transformation	No transformation
Replicate averaging	Averaging deactivated

Interpolation

Property	Value
Absolute potency [µg/ml]	0.09794
Potency of the working sample [µg/ml]	0.09794
Relative potency (percentage of reference concentration)	97.94%

Sample suitability tests

	Passed	Failed (rejected)	Failed (warning)	Info
Assay suitability tests	2	0	0	0
Sample suitability tests	1	0	0	0
Overall test result	Passed			

■ TST1 Test: Relative potency (percentage of reference concentration) Scope: Test samples only; Severity level: Warning		Passed
Calculated value	97.94450 80.00000 to 125.00000	